

# ACMG Standards and Guidelines for constitutional cytogenomic microarray analysis, including postnatal and prenatal applications: revision 2013

Sarah T. South, PhD<sup>1,2</sup>, Charles Lee, PhD<sup>3</sup>, Allen N. Lamb, PhD<sup>1,2</sup>, Anne W. Higgins, PhD<sup>4</sup> and Hutton M. Kearney, PhD<sup>5</sup>; for the Working Group for the American College of Medical Genetics and Genomics (ACMG) Laboratory Quality Assurance Committee

**Disclaimer:** These American College of Medical Genetics and Genomics *Standards and Guidelines* are developed primarily as an educational resource for clinical laboratory geneticists to help them provide quality clinical laboratory genetic services. Adherence to these standards and guidelines is voluntary and does not necessarily assure a successful medical outcome. These *Standards and Guidelines* should not be considered inclusive of all proper procedures and tests or exclusive of other procedures and tests that are reasonably directed to obtaining the same results. In determining the propriety of any specific procedure or test, the clinical laboratory geneticist should apply his or her own professional judgment to the specific circumstances presented by the individual patient or specimen. Clinical laboratory geneticists are encouraged to document in the patient's record the rationale for the use of a particular procedure or test, whether or not it is in conformance with these *Standards and Guidelines*. They also are advised to take notice of the date any particular guideline was adopted, and to consider other relevant medical and scientific information that becomes available after that date. It also would be prudent to consider whether intellectual property interests may restrict the performance of certain tests and other procedures.

Microarray methodologies, including array comparative genomic hybridization and single-nucleotide polymorphism–detecting arrays, are accepted as an appropriate first-tier test for the evaluation of imbalances associated with intellectual disability, autism, and multiple congenital anomalies. This technology also has applicability in prenatal specimens. To assist clinical laboratories in validation of

microarray methodologies for constitutional applications, the American College of Medical Genetics and Genomics has produced the following revised professional standards and guidelines.

*Genet Med* advance online publication 26 September 2013

**Key Words:** constitutional; guidelines; microarray; postnatal; prenatal; standards

## GENERAL CONSIDERATIONS

### Purpose of cytogenomic microarrays

Constitutional cytogenetic abnormalities include aneuploidy (extra or missing chromosomes) and structural aberrations (chromosomal gains and losses, translocations, inversions, insertions, and marker chromosomes). The cytogenomic microarray (CMA) platforms discussed in this guideline are those designed for the detection of DNA copy number gains and losses associated with unbalanced chromosomal aberrations. Regions with an absence of heterozygosity (AOH), also referred to as loss of heterozygosity, regions/runs of homozygosity, or long continuous stretches of homozygosity, may also be detected by platforms with single-nucleotide polymorphism (SNP)–detecting probes. Some regions with AOH may be indicative of uniparental isodisomy or regions of the genome identical by descent.

The utility of this technology for detection of gains and losses in patients with intellectual disabilities, autism, and/or congenital anomalies has been well documented, and CMA is now recommended as a first-tier test for these indications.<sup>1,2</sup>

### Advantages of CMAs

The benefits from the use of CMAs for detection of gains and losses of genomic DNA include:

1. Ability to analyze DNA from nearly any tissue, including archived tissue or tissue that cannot be cultured.
2. Detection of abnormalities that are cytogenetically cryptic by standard G-banded chromosome analysis.
3. Ability to customize the platform to concentrate probes in areas of interest.
4. Better definition and characterization of abnormalities detected by a standard chromosome study.
5. Interpretation of objective data, rather than a subjective visual assessment of band intensities.
6. Ability to detect copy neutral AOH with platforms incorporating SNP probes.
7. A ready interface of the data with genome browsers and databases.

<sup>1</sup>ARUP Laboratories, Salt Lake City, Utah, USA; <sup>2</sup>Department of Pathology, University of Utah, Salt Lake City, Utah, USA; <sup>3</sup>Brigham and Women's Hospital, Harvard Medical School, Boston, Massachusetts, USA; <sup>4</sup>Department of Pathology, UMass Memorial Medical Center and University of Massachusetts Medical School, Worcester, Massachusetts, USA; <sup>5</sup>Fullerton Genetics Center, Mission Health, Asheville, North Carolina, USA. Correspondence: Sarah T. South ([Sarah.South@aruplab.com](mailto:Sarah.South@aruplab.com))

Submitted 17 July 2013; accepted 17 July 2013; advance online publication 26 September 2013. doi:10.1038/gim.2013.129

**Limitations of CMAs**

Limitations of the use of CMAs include:

1. For most platforms, the inability to detect genetic events that do not affect the relative copy number of DNA sequences, e.g., molecularly balanced chromosomal rearrangements. However, CMAs may reveal copy number changes in apparently “balanced” chromosomal rearrangements, i.e., gains or losses, at or near the chromosomal breakpoint sites.
2. Low-level mosaicism for unbalanced rearrangements and aneuploidy may not be detected by CMAs. The sensitivity of the microarray for detection of mosaicism will be influenced by the platform, sample type, copy number state, DNA quality, data quality, and size of imbalance.
3. The chromosomal mechanism of a genetic imbalance may not be elucidated.
4. Tetraploidy or other ploidy levels may not be detected or may be difficult to detect.
5. Copy number variations (CNVs) of genomic regions not represented on the platform will not be detected.
6. Current CMA technologies are not designed to detect duplications and deletions below the level of detection according to probe coverage and performance, point mutations, gene expression, and methylation anomalies that may contribute to the patient’s phenotype.
7. No microarray platform will detect all mutations associated with a given syndrome. Therefore, it must be understood that failure to detect a copy number alteration at any locus does not exclude the diagnosis of a disorder associated with that locus.

**Microarray platform design and manufacture**

Different types of CMA platforms are currently available for clinical testing. The probes for these platforms may use either bacterial artificial chromosome–based DNA or oligonucleotide-based DNA. The oligonucleotide-based DNA may be designed to detect only a copy number alteration of a sequence as compared with a control, or may also be able to determine a specific genotype (or allele) associated with the probe (a SNP-detecting probe). The copy number of a probe may be determined either through a directly competitive hybridization of differentially labeled patient and control DNA or a comparative hybridization of the labeled patient DNA to an *in silico* reference set. The copy number data are graphed as a  $\log_2$  ratio of the probe intensities, with the expected normalized value equaling “0” (generally associated with two copies of genomic sequence), relative DNA gains having signals of greater intensity ( $\log_2 > 0$ ), and relative DNA losses having less intensity ( $\log_2 < 0$ ). For platforms with SNP-based probes, the copy number alteration should also correlate with the allelic information assuming sufficient coverage of the copy number alteration with SNP-detecting probes. For example, a region present in one copy should only have single SNP alleles identified in the region.

Microarray platform designs may have probes (i) targeted to specific regions of the genome for detection of imbalances known to be associated with congenital anomalies or neurocognitive impairments, (ii) distributed in a genome-wide manner with a specified distribution and spacing, or (iii) placed in both a targeted and genome-wide manner with varying distribution and spacing of probes for specific genomic regions as well as across the genome. The functional resolution of an array will be determined by both the intermarker probe spacing and the number of consecutive probes necessary to confidently identify a true CNV. The functional resolution may be different across different regions of the genome for a given platform due to probe density and may be different for a single copy number gain (two to three copies) versus a single copy number loss (two copies to one copy) of a DNA segment.

The American College of Medical Genetics and Genomics has published specific recommendations for the design and manufacture of CMA platforms.<sup>3</sup> At a minimum, for whole-genome platforms, the design should allow for detection of both gains and losses of 400 kb or larger, genome-wide, with exceptions to this minimal size resolution as necessary due to features of genomic architecture such as segmental duplication-rich regions. It is also desirable to have enrichment of probes targeting dosage-sensitive regions or genes well associated with congenital anomalies or neurocognitive impairments.

All probe descriptions/content and annotations should be openly accessible to the performing laboratory (see also “Annotation/databases” section). Details regarding the microarray design, the synthesis verification, and all quality control (QC) steps taken to validate and assess the performance and reproducibility of the array should be documented and provided by the manufacturer.

**FAMILIARIZATION WITH A NEW TECHNOLOGY FOR THE LABORATORY BEFORE VALIDATION**

The laboratory with little or no experience with microarray technology should become familiar with all aspects of the new technology before beginning the validation process, regardless of the regulatory status of the array. Familiarization begins with understanding the processes, features, and capabilities of the technology selected. The laboratory should gain experience with the instrumentation, platform design, software, reagents, methodology, technological limitations, workflows, DNA quality parameters, etc., by experimental sample runs. Similarly, the laboratory should become familiar with the features of each sample type the laboratory will process, as different sample types may have unique considerations for microarray data quality and clinical applicability. The laboratory must also be familiar with the potential imbalances and rearrangements associated with the clinical indications.

The use of samples well characterized as “normal” and “abnormal” by another method is valuable during the familiarization process to gain experience in the recognition of CNVs that may represent true biological variation or a probe/platform performance issue. It is suggested that laboratories use a

combination of data from well-characterized cases processed and run on their platform(s), data from other laboratories, and/or data available from online databases to gain and broaden their experience. Data sharing should involve a spectrum of array results and data quality.

Laboratories need to be able to recognize nonperforming (or nonresponsive) probes, technically induced artifacts, and other issues affecting data quality. Laboratories should become familiar with CNVs that are benign and/or common and resources to aid in the recognition and interpretation of CNVs.<sup>3-8</sup>

## VERIFICATION AND VALIDATION

### Verification of a Food and Drug Administration–approved/cleared platform

At the time of the publication of these guidelines, there are no commercially available Food and Drug Administration (FDA)-approved or FDA-cleared microarrays for this application. However, laboratories are advised to keep abreast of new developments in this rapidly developing technology.

For any FDA-approved or FDA-cleared microarrays where the laboratory plans to claim the test as FDA-approved/cleared, the approved protocol and intended use (usually included in the package insert) must be followed. The laboratory must verify that it can obtain comparable performance specifications as those established by the manufacturer with regard to accuracy, precision, and reportable range of results.

At the onset of verification, pass/fail criteria for the verification protocol should be established. If the prespecified acceptance criteria are not met, and a repeat or evaluation of the reasons for the failure does not resolve the concern, the laboratory should consider whether or not the array is appropriate for clinical testing.

Accuracy testing will measure the ability of the platform and software to detect known abnormalities. The accuracy evaluation is accomplished by running a series of previously characterized abnormal cases (this may be accomplished through sharing samples with an established laboratory). A minimum of 15 cases is recommended. To the extent possible, the laboratory should use abnormal samples that represent abnormalities that the array is designed to detect. This evaluation should include both a comparison of the findings from the region(s) expected to be abnormal as well as a comparison of the rest of the genome analyzed by the platform. The laboratory must document the concordance of the expected results and any unexpected findings. Because this technology may detect true alterations not previously identified, any unexpected findings that fall within the determined reportable range (as defined in “Validation of a new platform for the laboratory” section) should be further investigated to determine whether the finding represents true biological variation. This will involve the use of an alternative technology or microarray platform for correlation of the unexpected finding.

Precision testing should measure the reproducibility of repeated tests for the same result. The precision of the platform is established by running a minimum of two abnormal

## ACMG STANDARDS AND GUIDELINES

samples, each run multiple times in separate experiments. The concordance of the repeated runs should be documented, and any alterations should be considered (variability of breakpoints, calls, and potential reasons for variation, i.e., segmental duplication-rich region) as they pertain to the reportable range, functional resolution, and potential variability around breakpoints. Some variability around breakpoints may be expected due to segmental duplications and individual probe performance. Precision testing can allow for an assessment of breakpoints and potential impact on the clinical interpretation. Breakpoint variability that does not alter the clinical interpretation would be less concerning than variability that does alter the interpretation. Samples with multiple abnormalities are preferable as they maximize the number of findings for the precision study.

Any modification to the FDA-approved use of the product (as specified in the package insert) will be considered as off-label use, and therefore the microarray must then be validated as a non-FDA-approved platform.

### Validation of a non-FDA-approved platform

All platforms intended for clinical testing must be either FDA-approved/cleared and verified or must be validated by the performing laboratory. Validation is the process by which the laboratory measures the efficacy of the test in question by determining its performance characteristics when used as intended. This is necessary to demonstrate that it performs as expected and achieves the intended result. Validation is required when using laboratory-developed tests or modified FDA tests. The method and scope of the validation must be documented.

At the onset of validation, pass/fail criteria for the validation protocol should be established. If the prespecified acceptance criteria were not met, and a repeat or evaluation of the reasons for the failure does not resolve the concern, the laboratory should consider whether or not the array is appropriate for clinical testing.

The extent of work necessary for a validation can depend in part on whether the laboratory is validating a new microarray platform for the laboratory, validating a modified design of a previously validated platform, or adding additional sample types or intended uses to a previously validated platform. A new platform is defined as any new methodology or array introduced to the laboratory. A single microarray vendor may produce multiple similar platforms, but each must be validated independently. A modified design may include either minor modification to probe coverage, either through manufacturing of the array or by *in silico* probe filtering.

### Validation of a new microarray platform for the laboratory

Through the validation process, the laboratory must establish the performance characteristics of the microarray platform and accompanying software. The performance characteristics that must be established include the accuracy and precision of results, the analytical sensitivity and specificity, and the reportable ranges. Validations should be documented for each array

platform used for clinical testing, regardless of whether the laboratory has prior experience with a different platform.

The reportable range of results includes criteria to identify a CNV and criteria to report a CNV. Laboratories, with consideration of the manufacturer's recommendations, should identify the parameters specific to their platform (number of consecutive probes,  $\log_2$  ratios, SNP allele ratios, QC metrics, etc.) that are necessary to conclude that a copy number call represents a true CNV. As the functional resolution is a combination of probe density and number of probes necessary to identify a true CNV, the reportable range should be at or above the functional resolution of the platform. The reportable range should be determined before the evaluation of the validation set, and data from the familiarization process should be utilized. The reportable range may exclude well-characterized benign CNVs. If the reportable range is altered for the laboratory, the validation data should be re-evaluated with the new reportable range. However, if the previously identified validation samples do not contain abnormalities that challenge the altered reportable range, additional samples should be evaluated.

The accuracy evaluation is accomplished by running a minimum of 30 previously characterized abnormal controls. To the extent possible, the laboratory should use abnormal controls that represent abnormalities that the array is designed to detect. This should include both autosomal and sex chromosome abnormalities as duplications and deletions on the sex chromosomes may behave differently in each sex. Furthermore, blinding the evaluators to the expected abnormalities has the additional benefit of validating the settings, evaluation of data, and reportable range. Samples used for validation should represent a variety of findings with various sizes of abnormalities, combinations of gains and losses, various regions of the genome, and some aberrations that challenge the technical limits of detection for reportable DNA gains and losses.

Sample exchanges with a laboratory that is proficient with a similar microarray platform can provide a good source of samples for validation. Exchange of validated data sets (e.g., array files) between laboratories is recommended for additional experience in data analysis.

This evaluation should initially include a full review of the data to identify aberrations that meet the reportable range while blinded to the expected abnormality (as would fit the clinical workflow), followed by an unblinded comparison of the findings from the region(s) expected to be abnormal, as well as an evaluation of the rest of the genome analyzed by the platform. An evaluation of the regions expected to be normal is also important in assessing the probe behavior across the genome. The laboratory must document the concordance of the expected results and any unexpected findings. Evaluation should also include breakpoint evaluation with regard to gene content and genomic architecture. The laboratory should also recognize nonresponsive probes in a region expected to show loss or gain (this may be due to either poor performance probes or underlying genomic architecture). As this technology may detect true alterations not previously identified, any unexpected

CNVs that fall within your laboratory-determined reportable range should be further investigated to determine whether the finding represents true biological variation. This may involve the use of an alternative technology or microarray platform for correlation of the unexpected finding. As both expected and unexpected findings are evaluated, careful selection of the 30 samples is important and the ability to evaluate unexpected findings in the 30 samples should be considered.

Sensitivity and specificity are determined by the number of true-positive, true-negative, false-positive, and false-negative results in a validation data set that meet reporting criteria. However, for a whole-genome assay, all true positives and true negatives are not known. Therefore, specificity and sensitivity for genome-wide array platforms cannot be calculated as traditionally defined.

Sensitivity is evaluated by comparison of expected versus observed abnormalities, and this is then extrapolated to the rest of the genome. Rather than a traditional calculation of specificity, an evaluation of the positive predictive value of the assay is desirable. Determination of the positive predictive value will involve the identification of copy number calls that fall within the laboratory's determined reportable range and a determination of the proportion of those calls that are true. To improve the specificity of the platform, if certain probes are recognized to repeatedly act as false positives, these probes should be removed from future analyses. The identification of false-positive probes may be due to technical or biological variables considering that not all regions of the genome are amenable to accurate locus-specific evaluation of copy number with this technology. If probe content is masked by the laboratory, these changes should be documented. If the changes are sufficient to alter the performance of the platform, an evaluation of the validation data with the altered probe content is required.

The precision testing should measure the closeness of repeated test results to one another. The precision of the platform is established by running a minimum of two abnormal samples, each run multiple times in separate experiments. The concordance of the repeated runs should be documented, and any alterations should be considered (variability of breakpoints, calls, and potential reasons for variation, i.e., segmental duplication-rich region) as they pertain to the reportable range, functional resolution, and potential variability around breakpoints. Some variability around breakpoints may be expected due to genomic architecture and individual probe performance. The precision testing can allow for an assessment of breakpoints and potential impact on the clinical interpretation. Breakpoint variability that does not alter the clinical interpretation would be less concerning than variability that does alter the interpretation. Samples with multiple abnormalities are preferable as they maximize the number of findings for the precision study.

### **Validation of a new version of a previously established platform**

The definition of a new version should be limited to those situations in which a minimal number of probes are removed,

added, and/or replaced for the purpose of improved performance, and/or coverage is enhanced over a limited number of genomic regions. This would likely involve <10% of the total probe coverage, with no more than 5% probe removal. It should be recognized that these types of changes to an established platform are likely a rare event and most changes in platforms will require a full validation.

In the laboratory that is proficient with microarray technologies, a new version of a platform in use by the laboratory from the same manufacturer should be validated with a minimum of five abnormal samples.

Known abnormal samples from the previous version should be run using the new platform version for comparison to ensure that the performance meets the laboratory standards and to assess performance of probes added into higher-resolution platforms.

New content on an upgraded version should be assessed, if possible, using known abnormal sample(s) with variation in the region of the new content to determine performance.

The evaluation of this validation set of at least five samples should include data analyzed to determine whether the platform and software detected the expected abnormality. If other abnormalities are detected that meet the laboratory-reporting range, the validation should determine whether the findings represent true biological variation.

#### Validation of additional sample types on an established platform

It is expected that the initial validation will involve the most common sample type for the expected intended use. For example, if the intended use is postnatal evaluation, the sample type may be DNA extracted from peripheral blood. Because the quality of the DNA may vary from alternative tissue sources and this may add interference factors to the microarray analysis, use of DNA from alternative sample types requires an evaluation of the potential for interference.<sup>9</sup>

For a new sample type, an evaluation of the impact of the new sample type on data quality is necessary. The DNA extraction process should be part of the validation process. If there will be minimal changes to the processing or analysis, then a validation of the new sample type can involve equivalency of data quality with the new sample type. If alterations are made in the processing of the array or analysis (e.g., change of reference set), then a new validation is required. In addition, if the new sample type requires a different reportable range, then a new validation is required.

#### Validation of the allelic differentiation potential of SNP-detecting platforms

The detection of AOH is not in and of itself diagnostic but can identify a concern that would require additional testing such as sequence-based mutation analysis or uniparental disomy testing. However, as AOH may be reported by the laboratory, evaluation of the performance of the SNP-detecting probes to define regions of AOH should be included in the validation.

Reports of these findings must clearly state that the finding is not diagnostic.

Given sufficient probe density, there should be a correlation between copy number state and SNP allele state.

A minimum of five samples need to contain expected copy neutral AOH in addition to CNVs. Interlaboratory comparisons of samples that contain known uniparental isodisomy or regions identical by descent are recommended. This comparison should address the data types that would be included in a report, such as approximate regions of AOH and approximate percentage of genome identical by descent. The detection and accurate size assessment of AOH by SNP-based arrays depend on the density of SNP probes. If the validation method does not address accuracy of breakpoints in AOH calls, reports should reflect this uncertainty. Inaccurate size estimation for regions of AOH could lead to unwarranted follow-up testing for uniparental isodisomy, somatic loss of heterozygosity, and/or autozygosity mapping.<sup>10</sup>

#### Mosaicism detection

Low-level mosaicism for unbalanced rearrangements and aneuploidy may not be detected by microarray analysis. In addition, the level of detectable mosaicism will vary by size, region of genome, copy number state, DNA quality and data quality. Therefore, it is not likely that a specific level of mosaicism can always be identified uniformly throughout the genome, and this limitation should be recognized.

Without extensive validation to determine specific levels of mosaic detection for a wide variety of CNV sizes and genomic regions, it is not recommended that this technology be used to rule out mosaicism. However, experience in mosaicism identification is desirable to maximize opportunity for detection. Methods for determining detectable levels of mosaicism include dilution studies and analysis of the sample by other quantitative methods. Fluorescence *in situ* hybridization analysis of fresh (uncultured) samples provides a reliable means to establish the level(s) of mosaicism detectable by microarray. Conventional cytogenetic analysis of metaphase cells provides information about mosaicism but may not accurately reflect levels of mosaicism. The laboratory director should determine the method used by the laboratory. More than one method is recommended.

For cells in suspension, dilution studies using samples with known CNVs may help to determine detectable levels of mosaicism. This method can provide an effective means to establish thresholds; however, it may have limitations as an artificial method. Dilution studies for SNP-detecting arrays may not be possible because they may introduce additional genotypes that complicate the analysis.

The detection of mosaicism may include information from both the log<sub>2</sub> ratio and the SNP allele pattern as applicable for each platform.

Be aware that microarray analysis gives a relative level of copy number across the cells within the sample but does not provide a cell-by-cell determination of copy number (e.g., trisomy in 60% vs. tetrasomy in 30% of cells).

**Special considerations for validation of prenatal specimens**

Experience with postnatal arrays and with common and rare CNVs is important for the processing and interpretation of array results for prenatal specimens. For validation, a distinction should be made between cultured amniocytes and chorionic villus sampling (CVS) cells and uncultured (direct) amniocytes and villi. The validation performed depends on whether the platform has been previously validated for postnatal use or is new to the laboratory and whether both cultured and uncultured cells will be used.

For cultured amniocytes and CVS cells, if prenatal array analysis is performed on an array platform new to the laboratory, the issues and process discussed in the “Validation of a new platform for the laboratory” section apply, and a minimum of 30 previously characterized cases should be processed. Due to the difficulty of obtaining abnormal prenatal specimens, this collection of 30 samples will likely include some previously characterized as normal cases. Therefore, additional experience with abnormal array findings through additional tissue types and data exchanges should occur, to ensure that a wide variety of abnormalities have been evaluated both in-house and *in silico*.

For a previously validated platform for postnatal use, the addition of prenatal specimens requires an understanding of the potential issues that these samples can present regarding data quality. The DNA extraction process should be part of the validation process. If the laboratory will perform analysis on cultured amniocytes and CVS, both sample types should be represented in the validation.

**Prenatal samples (including products of conception).** Healthy cultures established from amniocytes, villi, and fetal tissue yield an adequate quantity and quality of DNA and can be viewed as essentially equivalent for validation purposes. However, the laboratory should be aware of factors that can affect DNA yield and data quality including culture age, growth rate, confluency, and shipping conditions.

Because uncultured cells may yield different amounts and quality of DNA, additional validation is required to become familiar with potential differences as compared with cultured cells. Parameters to consider for uncultured amniocytes include method of DNA extraction, volume, and gestational age given that these parameters influence the amount and quality of DNA. For example, uncultured amniocytes yield less DNA than cultured cells; however, the quality of the DNA is generally higher from uncultured cells.

Because villi represent a more complicated tissue with different cell types/layers (syncytiotrophoblast, cytotrophoblast, and mesenchymal core), DNA may be extracted from all cell types, or the laboratory may eliminate or concentrate on different cell layers for DNA extraction.<sup>11</sup>

**Special quality assurance requirements for prenatal specimens.** Back-up cultures of all prenatal samples undergoing array analysis should be established and maintained for the purposes of (i) possible array failures on direct extractions, (ii)

evaluation of possible mosaicism on an independent culture, and (iii) the need to perform metaphase chromosome or fluorescence in situ hybridization analysis to investigate CNVs.

Maternal cell contamination (MCC) analysis should be performed on all prenatal samples, unless contamination is otherwise excluded. MCC can result from direct samples of amniocytes with blood admixture, CVS samples not successfully cleaned of maternal decidua (a more frequent problem with products of conception), and cell cultures undergoing extensive subculturing resulting in expansion of maternal cells. When undetected, MCC can result in missed detection or misinterpretation of copy number changes, even in the context of a male result. Laboratories should understand that MCC can be detected by array software (i.e., SNP-based platforms) or, in the case of male fetuses, by a shift of the sex chromosome plots (mimicking mosaicism). Laboratories should also understand how the presence of MCC can affect detection of CNVs, including different types (gains and losses) and different sizes (small versus large gains and losses).

Mosaicism may be detected in prenatal samples and may represent culture artifact (pseudomosaicism), true fetal mosaicism, or, for CVS, confined placental mosaicism. Careful investigation may be required to determine the fetal genotype. For traditional chromosome analysis, algorithms have been developed to deal with confined placental mosaicism and pseudomosaicism. These algorithms will also need to be developed for microarray analysis and will depend on whether the analysis used direct or cultured cells, and if the mosaicism can be confirmed on an independent culture.

**ESTABLISHING A REFERENCE DNA SET**

Depending on the platform used, the reference DNA set may come from a single individual or multiple individuals and may be sex matched or mismatched, and may be used *in silico* or as a direct competitive hybridization. The laboratory should understand the benefits and limitations of each scenario. The laboratory should also consider how the data quality is affected by the source and components of the reference DNA set. For example, data quality is likely improved when the conditions used for data acquisition from the reference set closely match the experimental conditions used for the test.<sup>12</sup> Any changes to the reference DNA set require a verification of the quality and accuracy of results obtained with the new reference DNA set as compared with the previous reference DNA set, especially because changes to the control can result in variation of results, particularly within polymorphic regions. For arrays that use *in silico* controls, versioning should be documented.

**SOFTWARE CONSIDERATIONS**

The laboratory should recognize software limitations and the need for manual and visual inspection of the data for aberration and mosaicism detection because the software may not flag all relevant calls that may be identifiable by a visual inspection of the data. To verify that the method for result generation (including software calls and manual inspection) detects known

aberrations at certain mosaicism levels, the system should be challenged with different types of aberrations. During the familiarization phase, the settings should be explored and optimized for aberration detection and then established parameters should be used consistently throughout the validation process. Algorithmic parameter settings may be different for various sample types.

Changes to the software settings from those used during the validation may require a re-analysis of at least a subset of the validation data using the new settings to identify any changes to the performance characteristics of the microarray platform. Such changes may include, but are not limited to, new annotation libraries, changes to any *in silico* reference set, or any changes to the aberration-calling algorithm.

The laboratory should understand that most normalization algorithms assume a primarily diploid state, which may obscure the detection of polyploidy. In some situations, the allele pattern may assist with the detection of triploidy but may not identify tetraploidy. Both situations are likely rare in the postnatal constitutional population but may be present as mosaic findings as well as in prenatal settings.

The laboratory should document the software, parameters, and rules used in the analysis of the microarray, as well as all limitations of the analysis program.

## QUALITY CONTROL

### Identification

For each array, the slide identification number, sample sex, control sex, and sample-tracking control (if applicable) should be verified. Discrepancies in the documentation from the physical sample should be investigated and resolved before processing.

### DNA requirements

The laboratory should establish the minimum DNA requirements to perform testing. Each laboratory should have established parameters for the determination of the sample quality and quantity and criteria for adequacy of each. If a sample does not meet these minimum requirements and is deemed suboptimal, the recommended action is to reject the specimen and request a repeat specimen.

### Equipment calibration, maintenance, and QC

Equipment, instrumentation, and methodologies employed during the validation and use of microarray platforms should be calibrated, monitored for QC, and regularly maintained as appropriate. Quality metrics should be established whenever possible throughout the assay. Laboratories should ensure that documentation and safeguards are provided by the software manufacturer and that data are processed and summarized in a consistent fashion for every clinical analysis. Most analysis software provides a hierarchy of users with customizable permissions, which enable the laboratory director or supervisor to prevent modification of analysis settings so that all specimens are analyzed consistently. Any changes to data processing should be validated and documented.

## ACMG STANDARDS AND GUIDELINES

### General QC metrics

Every microarray platform has defined quality metric values, e.g., adequate dye incorporation and/or amplification, fluorescence intensities, signal-to-background noise ratio, and standard deviation or standard error. Standard cutoff values and acceptable limits should be established for these metrics to ensure that the generated results are reliable and precise enough to be used for a clinical assessment. Quality metrics should be monitored for DNA labeling, hybridization efficiency, data generation and analysis, and other platform-specific parameters. The QC metrics should be incorporated into the laboratory quality assurance and quality improvement programs to monitor analytical variables.

### Data quality

The quality of the data will affect the ability to detect genomic aberrations. Therefore, it is absolutely necessary for the laboratory to understand the within-array metrics provided by the analysis software and how each of these metrics reflects the quality of the data. There are a number of metrics that provide a measurement of signal to noise (i.e., artifactual random variance unrelated to genomic location) in the data, such as the difference between the log ratio values of consecutive probes and the spread of the derivative log ratio values after outlier rejection. Similar metrics of variance exist for each platform.

The laboratory should establish acceptable ranges for each QC metric chosen to represent data quality by the laboratory. These ranges are often provided by the manufacturer. However, the laboratory may want to modify these ranges on the basis of their experience with the arrays during the validation process. The laboratory should establish criteria for next steps, should the data fall outside of these established ranges.

### Annotation/databases

An integral part of the data analysis is accessibility and use of private and public annotations/databases during the analysis process. Because these annotations are critical for interpretation, it is important that these tools are carefully constructed and applied by the laboratory or software manufacturer. All critical annotations should be thoroughly vetted, and the source(s) should be verified. For all reportable calls, the genomic content should be verified by an independent database source. The manufacturers should provide mechanism(s) for updates to these annotations. Documentation of resources and databases accessed for interpretation is recommended.

### Verification of new lots of microarrays/reagents

Verification should ensure that new lots of microarray slides and/or reagents perform in the same manner as the previous lot. The manufacturer should supply documentation of the QC comparison between lots (e.g., oligonucleotide synthesis verification, accuracy of SNP calls or other defined control parameters). New lots of reagents (e.g., new labeling kits, consumables) should have documented equivalency between runs. This may be accomplished by documenting that the QC metrics meet certain set parameters for the new lot of reagents.

## Confirmation of specific CNVs

With proper technical performance and analytical validation, it should not be necessary for the performing laboratory to further confirm a CNV called with the laboratory-validated parameters, after the validation stage. Each laboratory should establish a threshold (number of probes and/or genomic size, as well as other QC metrics) for declaring what constitutes a reportable abnormality with their assay. Features to keep in mind when assessing copy number changes are the appropriate log ratio difference between data, the presence of uniform contiguous probe behavior within and adjacent to call, sharp copy number state transitions at breakpoint boundaries, supportive SNP allele states (when applicable), and evaluation of least processed  $\log_2$  ratio data (e.g., weighted versus not weighted). Any call-specific quality score provided by the software may be considered.

Given that it is desirable to maximize detection of aberrations of clinically important genes and of aberrations in mosaic form (which may not generate a robust copy number call), it is acceptable and appropriate at the discretion of the performing laboratory to evaluate calls that do not meet the laboratory-validated parameters. These calls may be flagged for review and correlated with the patient's clinical indication, but should be confirmed by an independent methodology if reported.

## USE OF ALTERNATIVE TECHNOLOGIES FOR MECHANISM DETERMINATION

Determination of the mechanism leading to the detected CNV may be considered on a case-by-case basis because this may lead to better determination of recurrence risk. Some mechanisms can be identified through the combination of both the CNV and recognition of the genomic location of the altered material, or the genomic structure surrounding the alteration. Examples include both terminal and insertional translocations and ring or marker chromosomes. The appropriate alternative technology may depend on the size, type, and location of the identified CNV and the likely mechanism of formation. Therefore, use of these alternative technologies should be considered as separate testing and should use validated technologies performed and interpreted by appropriately trained personnel.

## INTERPRETATION AND REPORTING

For further guidance on interpretation and reporting, refer to recently published guidelines from the American College of Medical Genetics and Genomics for interpretation and reporting of postnatal constitutional copy number variants<sup>13</sup> and for reporting suspected consanguinity as an incidental finding of genomic testing.<sup>14</sup>

## METHODOLOGY AND DISCLAIMERS

All reports should include a brief description of methodology, including platform specifics and reporting criteria. Disclaimers should be included as appropriate and required.

## Example: testing limitations

Current microarray analysis technologies will detect only gains and losses of genomic segments. Therefore, a normal microarray result does not exclude mutations (nucleotide base-pair changes) in any gene represented on the microarray, gains and losses below the level of resolution of the platform, a balanced rearrangement, or epigenetic events. Additional testing may be appropriate for certain syndromes or conditions when the microarray analysis yields normal results.

## Alternative example

This microarray platform will not detect truly balanced chromosomal rearrangements, point mutations, or imbalances of regions not represented on the microarray, and may not detect mosaicism. Failure to detect an alteration at any locus does not exclude all anomalies at that locus.

## Example: disclaimer for a non-FDA-approved microarray platform

This test was developed and its performance characteristics determined by (your laboratory name here) as required by Clinical Laboratory Improvement Act (CLIA) 1988 regulations. It has not been cleared or approved for specific uses by the US Food and Drug Administration. Pursuant to the 1988 CLIA requirements, this laboratory has established and verified the test's accuracy and precision.

## PROFICIENCY TESTING

The laboratory should participate in an external proficiency testing program through an appropriate deemed organization (e.g., the College of American Pathologists). The laboratory should also establish internal proficiency testing of normal and abnormal samples as part of the laboratory internal quality assurance program and ongoing quality improvement program. Correlation between microarray results run in parallel on different array platforms or correlation of microarray results with conventional cytogenetic and/or fluorescence *in situ* hybridization results may be sufficient to provide ongoing proficiency. Proficiency testing should be performed according to 1988 CLIA guidelines.

Documentation of participation and the performance results of all internal and external proficiency tests must be retained by the laboratory and made available to all accreditation agency inspectors.

## LABORATORY ACCREDITATION AND PERSONNEL QUALIFICATIONS

Laboratory personnel must have documentation of education, degrees, and certifications as appropriate for level of testing, as well as training, competency assessments, and continuing education as required by appropriate regulatory bodies, e.g., College of American Pathologists, CLIA, Centers for Medicare & Medicaid Services. The testing laboratory must have CLIA certification and state certifications as required to provide clinical testing. College of American Pathologists accreditation is strongly encouraged.

## RETENTION OF FILES AND DOCUMENTATION

Laboratories should make explicit in their policies which file types and what length of time each type will be retained and that data retention policy must be in accordance with local, state, and federal requirements. CLIA regulations (Sec. 493.1105) require storage of analytic systems records and test reports for at least 2 years. For more specific suggestions for microarray technologies, we recommend that the laboratory consider a minimum of 2-year storage of a file type that would allow regeneration of the primary results as well as re-analysis with improved analytic pipelines. In addition, laboratories should consider retention of the aberrations identified in the analysis, along with the final clinical test report interpreting the subset of clinically relevant variants, for as long as possible, given the likelihood of a future request for reinterpretation of variant significance.

## CONCLUSIONS

Each new technological development in the field of genetics brings with it the desire to apply the technology to improve medical care. The transition of a new technology from the research bench into the clinical realm of diagnostic testing must be accompanied by extensive validation to ensure that the results reported to the health-care provider are accurate and reliable for use in patient-care decision making.

Microarray technologies provide a high-resolution view of the whole genome. Medical laboratory professionals must be prepared to identify, interpret, and report the results with clinical relevance, while keeping in mind the social, ethical, and legal responsibilities of reporting genetic information. The interpretation of the data from microarray analysis into clinically relevant information is a difficult and complex undertaking and is the practice of medicine. No algorithm for CNV interpretation can substitute for adequate training and knowledge in the field of genetics. We recommend that genomic microarray analysis be performed in laboratories overseen by individuals with appropriate professional training (American Board of Medical Genetics–certified clinical cytogeneticists or clinical molecular geneticists, or American Board of Medical Genetics/American Board of Pathology–certified molecular genetic pathologists) and that the interpretation and reporting of clinical genomic microarray findings be performed by these same certified individuals.

## ACKNOWLEDGMENTS

This study was approved by the American College of Medical Genetics and Genomics Board of Directors, 10 July 2013.

## DISCLOSURE

All members of this working group are directors of clinical laboratories that use genomic microarray technologies (please see affiliations for locations of clinical laboratories). S.T.S. has received honoraria from Affymetrix, a manufacturer of genomic microarray platforms, for speaking engagements. Also, S.T.S. is a consultant to Lineagen, a provider of genomic microarray testing services. The other authors declare no conflict of interest.

## REFERENCES

- Manning M, Hudgins L; Professional Practice and Guidelines Committee. Array-based technology and recommendations for utilization in medical genetics practice for detection of chromosomal abnormalities. *Genet Med* 2010;12:742–745.
- Miller DT, Adam MP, Aradhya S, et al. Consensus statement: chromosomal microarray is a first-tier clinical diagnostic test for individuals with developmental disabilities or congenital anomalies. *Am J Hum Genet* 2010;86:749–764.
- Kearney HM, South ST, Wolff DJ, Lamb A, Hamosh A, Rao KW; Working Group of the American College of Medical Genetics. American College of Medical Genetics recommendations for the design and performance expectations for clinical genomic copy number microarrays intended for use in the postnatal setting for detection of constitutional abnormalities. *Genet Med* 2011;13:676–679.
- Lee C, lafrate AJ, Brothman AR. Copy number variations and clinical cytogenetic diagnosis of constitutional disorders. *Nat Genet* 2007;39(suppl 7):S48–S54.
- Firth HV, Richards SM, Bevan AP, et al. DECIPHER: Database of Chromosomal Imbalance and Phenotype in Humans Using Ensembl Resources. *Am J Hum Genet* 2009;84:524–533.
- Riggs ER, Church DM, Hanson K, et al. Towards an evidence-based process for the clinical interpretation of copy number variation. *Clin Genet* 2012;81:403–412.
- Sayers EW, Barrett T, Benson DA, et al. Database resources of the National Center for Biotechnology Information. *Nucleic Acids Res* 2012;40(Database issue):D13–D25.
- Zhang J, Feuk L, Duggan GE, Khaja R, Scherer SW. Development of bioinformatics resources for display and analysis of copy number and other structural variants in the human genome. *Cytogenet Genome Res* 2006;115:205–214.
- van Heesch S, Mokry M, Boskova V, et al. Systematic biases in DNA copy number originate from isolation procedures. *Genome Biol* 2013;14(4):R33.
- Mason-Suares H, Kim W, Grimmitt L, et al. Density matters: comparison of array platforms for detection of copy-number variation and copy-neutral abnormalities. *Genet Med* 2013;15:706–712.
- Cytogenetics TAFC. *Professional Guidelines for Clinical Cytogenetics and Clinical Molecular Genetics: QF-PCR for the Diagnosis of Aneuploidy, Best Practice Guidelines*. In: Genetics BSfH (ed), v3.01. Clinical Genetics Unit, Birmingham Women's Hospital; Birmingham, England, B 15 2TG: ACC Heads of Department; 2012:1–16.
- Pinto D, Darvishi K, Shi X, et al. Comprehensive assessment of array-based platforms and calling algorithms for detection of copy number variants. *Nat Biotechnol* 2011;29:512–520.
- Kearney HM, Thorland EC, Brown KK, Quintero-Rivera F, South ST; Working Group of the American College of Medical Genetics Laboratory Quality Assurance Committee. American College of Medical Genetics standards and guidelines for interpretation and reporting of postnatal constitutional copy number variants. *Genet Med* 2011;13:680–685.
- Rehder CW, David KL, Hirsch B, Toriello HV, Wilson CM, Kearney HM. American College of Medical Genetics and Genomics: standards and guidelines for documenting suspected consanguinity as an incidental finding of genomic testing. *Genet Med* 2013;15:150–152.